HEREFORD SCALE CO.
1 VAGA STREET,
HEREFORD. HR2 7AT.
TEL 01432 356472 FAX 01432 352646

CERTIFICATE OF CALIBRATION

CERT No

CUSTOMER ADDRESS

SAMPLE

TEST DATE	10th May 2007
NEXT DUE	May 2008
FREQUENCY	12 Months

CERTIFICATE NUMBERS

3574 3575 08491 08492 4427

LOCATION	MAKE	TYPE	SERIAL No.	CAPACITY
As Above	T-Scale	T28-6	840137011	6 kg
As Above	T-Scale	T28-6	840137010	6 kg
As Above	Uwe	MII-6000	R3071	6 kg
As Above	Uwe	MII-6000	R3073	6 kg

PAGE 1 of 1

Authorised Name. S J Beavan

Engineer's signature. _____

1. Certificate, Tolerances and Accuracy.
The certificate issued under this service is based on readings taken at a particular point in time and a particular location. Unless otherwise agreed, tolerances will be those permitted by current Trading Standards regulations.
2. Traceability.
The Test Weights are certified for conformity and traceability to National Standards. The appropriate certificates of conformity for the Test Weights used will be quoted on the certificate. Traceability which is a requirement for ISO9001 implies that there is an unbroken chain of acceptable , authorised certificates of conformity back to the National Physical Laboratory National Standards.
3. Hereford Scale Co, as a member of the UK Weighing Federation agree to abide by the UKWF Calibration Code of Practice.

HEREFORD SCALE CO.
1 VAGA STREET,
HEREFORD. HR2 7AT.
TEL 01432 356472 FAX 01432 352646

CERTIFICATE OF CALIBRATION

CERT No

CUSTOMER ADDRESS

SAMPLE

TEST DATE	7th August 2007
NEXT DUE	February 2008
FREQUENCY	6 Months
MAKE	Sartorius
CAPACITY	210g
DIVISIONS	0.00001g
MODEL	ME215S - OCE GENIUS
TOLERANCE	0g to 50g + / - 1mg
	50g to 200g + / - 2mg
SERIAL No	(LAB 59)
	200g + + / - 3mg

CERTIFICATE NUMBERS

05504

AS FOUND			DEFINITIVE	
TEST POINTS	READING	ERROR	READING	ERROR
Zero	0.00000g	0mg	0.00000g	0mg
0.00100g	0.00100	0	0.00100	0
0.01000g	0.01000	0	0.01000	0
1.00000g	1.00000	0	1.00000	0
10.00000g	10.00003	+ 0.03	10.00003	+ 0.03
50.00000g	50.00004	+ 0.04	50.00004	+ 0.04
100.00000g	99.99986	- 0.14	99.99986	- 0.14
200.00000g	199.99982	- 0.18	199.99982	- 0.18
REPEATABILITY			REPEATABILITY	
20.00000g	20.00005	+ 0.05	20.00005	+ 0.05
20.00000g	20.00004	+ 0.04	20.00004	+ 0.04
20.00000g	20.00002	+ 0.02	20.00002	+ 0.02
OFF-CENTRE LOADING			OFF-CENTRE LOADING	
Corner 1	20.00010	+ 0.10	20.00010	+ 0.10
Corner 2	20.00006	+ 0.06	20.00006	+ 0.06
Corner 3	20.00010	+ 0.10	20.00010	+ 0.10
Corner 4	20.00003	+ 0.03	20.00003	+ 0.03

Authorised signature. _____

1. Certificate, Tolerances and Accuracy.
The certificate issued under this service is based on readings taken at a particular point in time and a particular location. Unless otherwise agreed, tolerances will be those permitted by current Trading Standards regulations.
2. Traceability.
The Test Weights are certified for conformity and traceability to National Standards. The appropriate certificates of conformity for the Test Weights used will be quoted on the certificate. Traceability which is a requirement for ISO9001 implies that there is an unbroken chain of acceptable , authorised certificates of conformity back to the National Physical Laboratory National Standards.
3. Hereford Scale Co, as a member of the UK Weighing Federation agree to abide by the UKWF Calibration Code of Practice.